

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062918\
 Data File : VX003039.D
 Acq On : 29 Jun 2018 19:51
 Operator : JC/MD
 Sample : J3694-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-3

Quant Time: Jun 30 00:22:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274352	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	378903	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	362185	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	203206	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	178274	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
35) Dibromofluoromethane	5.51	113	148576	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
50) Toluene-d8	8.72	98	541496	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	11.14	95	190332	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
Target Compounds						
16) Acetone	2.45	43	5698	3.63	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062918\
Data File : VX003039.D
Acq On : 29 Jun 2018 19:51
Operator : JC/MD
Sample : J3694-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-3

Quant Time: Jun 30 00:22:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

