

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062918\
 Data File : VX003053.D
 Acq On : 30 Jun 2018 02:29
 Operator : JC/MD
 Sample : J3707-23 5X
 Misc : 6.12g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021SB489-25-27-180622

Quant Time: Jun 30 10:08:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	265585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	371409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	363727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	213318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176235	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
35) Dibromofluoromethane	5.51	113	141698	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
50) Toluene-d8	8.72	98	543594	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	11.14	95	196266	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
Target Compounds						
4) Vinyl Chloride	1.40	62	8774	2.92	ug/l	# 59
12) 1,1-Dichloroethene	2.36	96	1682	0.64	ug/l	90
27) cis-1,2-Dichloroethene	4.60	96	369356	115.47	ug/l	82
44) Trichloroethene	7.21	130	235046	66.62	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062918\
Data File : VX003053.D
Acq On : 30 Jun 2018 02:29
Operator : JC/MD
Sample : J3707-23 5X
Misc : 6.12a/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021SB489-25-27-180622

Quant Time: Jun 30 10:08:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

