

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062918\
 Data File : VX003061.D
 Acq On : 30 Jun 2018 06:02
 Operator : JC/MD
 Sample : J3675-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-0213

Quant Time: Jun 30 10:27:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	302643	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	422977	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	396993	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	230939	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	197591	47.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.48%	
35) Dibromofluoromethane	5.51	113	160590	47.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.88%	
50) Toluene-d8	8.72	98	587428	48.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.04%	
62) 4-Bromofluorobenzene	11.14	95	217477	46.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062918\
Data File : VX003061.D
Acq On : 30 Jun 2018 06:02
Operator : JC/MD
Sample : J3675-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S73-0213

Quant Time: Jun 30 10:27:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

