

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062918\
 Data File : VX003066.D
 Acq On : 30 Jun 2018 08:15
 Operator : JC/MD
 Sample : J3675-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-0217

Quant Time: Jun 30 09:35:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	389916	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	358544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	205942	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	181626	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
35) Dibromofluoromethane	5.51	113	148731	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	8.72	98	540661	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
62) 4-Bromofluorobenzene	11.14	95	191338	44.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	7.22	130	1094	0.30	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062918\
Data File : VX003066.D
Acq On : 30 Jun 2018 08:15
Operator : JC/MD
Sample : J3675-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S73-0217

Quant Time: Jun 30 09:35:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

