

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062920\
 Data File : VX017023.D
 Acq On : 29 Jun 2020 21:34
 Operator : JC/SP
 Sample : L3146-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 200624027-01-TRIP-BLANK

Quant Time: Jun 30 06:27:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	255155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	446675	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	457804	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	245173	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	166867	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
35) Dibromofluoromethane	5.47	113	144085	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
50) Toluene-d8	8.70	98	543092	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	11.12	95	227963	55.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.62%	
Target Compounds						
16) Acetone	2.42	43	247074	209.968	ug/l	100
19) Methyl tert-butyl Ether	3.17	73	10858	1.282	ug/l	97
20) Methylene Chloride	2.84	84	20392	6.881	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062920\
Data File : VX017023.D
Acq On : 29 Jun 2020 21:34
Operator : JC/SP
Sample : L3146-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
200624027-01-TRIP-BLANK

Quant Time: Jun 30 06:27:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

