

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062920\
 Data File : VX017028.D
 Acq On : 29 Jun 2020 23:28
 Operator : JC/SP
 Sample : L3097-02DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 25BR-20200623DL

Quant Time: Jun 30 06:45:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	247451	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	438162	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	443919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231761	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	165322	54.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.38%	
35) Dibromofluoromethane	5.47	113	141164	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
50) Toluene-d8	8.70	98	530245	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	11.12	95	218632	54.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.14%	
Target Compounds						
4) Vinyl Chloride	1.39	62	3985	1.482	ug/l	93
21) trans-1,2-Dichloroethene	3.14	96	1535	0.565	ug/l #	87
27) cis-1,2-Dichloroethene	4.57	96	112243	36.462	ug/l	90
44) Trichloroethene	7.19	130	213128	63.560	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062920\
Data File : VX017028.D
Acq On : 29 Jun 2020 23:28
Operator : JC/SP
Sample : L3097-02DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
25BR-20200623DL

Quant Time: Jun 30 06:45:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

