

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX022997.D
 Acq On : 29 Jun 2021 12:47
 Operator : JC/MD
 Sample : M2838-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB04-20210621

Quant Time: Jun 30 01:43:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	111506	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	95651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35221	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52044	53.169	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.340%
35) Dibromofluoromethane	5.397	113	36375	51.371	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.740%
50) Toluene-d8	8.653	98	142950	51.511	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.020%
62) 4-Bromofluorobenzene	11.085	95	47765	46.118	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
Data File : VX022997.D
Acq On : 29 Jun 2021 12:47
Operator : JC/MD
Sample : M2838-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04-20210621

Quant Time: Jun 30 01:43:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

