

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023000.D
 Acq On : 29 Jun 2021 13:57
 Operator : JC/MD
 Sample : M2875-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Jun 30 01:43:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53476	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111106	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	95730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50411	51.945	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.880%
35) Dibromofluoromethane	5.397	113	33746	47.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.660%
50) Toluene-d8	8.653	98	134005	48.462	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.920%
62) 4-Bromofluorobenzene	11.085	95	46010	44.584	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.160%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	3045	7.077	ug/l	95
20) Methylene Chloride	2.794	84	1685	2.332	ug/l	# 93
52) Toluene	8.720	92	1834	0.892	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
Data File : VX023000.D
Acq On : 29 Jun 2021 13:57
Operator : JC/MD
Sample : M2875-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Jun 30 01:43:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

