

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023002.D
 Acq On : 29 Jun 2021 14:44
 Operator : JC/MD
 Sample : M2875-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Jun 30 01:43:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53094	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.775	114	110013	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54926	57.004	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	114.000%		
35) Dibromofluoromethane	5.391	113	33932	48.571	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.140%		
50) Toluene-d8	8.653	98	134894	49.268	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.540%		
62) 4-Bromofluorobenzene	11.085	95	45785	44.806	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.620%		
Target Compounds						Qvalue
16) Acetone	2.386	43	1657	3.879	ug/l	# 83
30) Chloroform	5.098	83	3334	2.279	ug/l	99
40) Benzene	6.050	78	1205	0.358	ug/l	# 87
64) Tetrachloroethene	9.275	164	363	0.530	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
Data File : VX023002.D
Acq On : 29 Jun 2021 14:44
Operator : JC/MD
Sample : M2875-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Jun 30 01:43:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

