

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023011.D
 Acq On : 29 Jun 2021 18:17
 Operator : JC/MD
 Sample : M2875-19DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12DL

Manual Integrations
 APPROVED

MMDadoda
 7/2/2021 6:21:40 PM

Quant Time: Jun 30 01:45:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59614	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	120793	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	107165	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43071	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57241	52.909	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.820%			
35) Dibromofluoromethane	5.397	113	38508	50.202	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.400%			
50) Toluene-d8	8.653	98	152366	50.683	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.360%			
62) 4-Bromofluorobenzene	11.085	95	55924	49.845	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.680%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	1710	3.565	ug/l	96
31) Cyclohexane	5.483	56	5834	3.820	ug/l #	87
39) Methylcyclohexane	7.379	83	6456	4.314	ug/l #	78
40) Benzene	6.044	78	73786	19.944	ug/l	99
52) Toluene	8.726	92	5031	2.250	ug/l	91
67) Ethyl Benzene	10.195	91	103159	24.823	ug/l	94
68) m/p-Xylenes	10.305	106	157617	102.130	ug/l	91
69) o-Xylene	10.646	106	19669	13.095	ug/l	92
73) Isopropylbenzene	10.963	105	12836	3.522	ug/l	97
78) n-propylbenzene	11.305	91	26261	5.955	ug/l	95
80) 1,3,5-Trimethylbenzene	11.457	105	41445	13.849	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	163063	54.139	ug/l	97
85) sec-Butylbenzene	11.890	105	2003m	0.526	ug/l	
89) n-Butylbenzene	12.323	91	5235	1.818	ug/l #	28
95) Naphthalene	13.780	128	100823	29.535	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
Data File : VX023011.D
Acq On : 29 Jun 2021 18:17
Operator : JC/MD
Sample : M2875-19DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12DL

Manual Integrations
APPROVED

MMDadoda
7/2/2021 6:21:40 PM

Quant Time: Jun 30 01:45:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

