

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023012.D
 Acq On : 29 Jun 2021 18:40
 Operator : JC/MD
 Sample : M2889-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-2

Quant Time: Jun 30 01:45:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59155	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	119769	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55048	51.277	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.560%
35) Dibromofluoromethane	5.397	113	37553	49.375	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.760%
50) Toluene-d8	8.653	98	148364	49.773	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.085	95	53036	47.675	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.340%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2395	5.032	ug/l	# 86
67) Ethyl Benzene	10.201	91	1209	0.294	ug/l	91
68) m/p-Xylenes	10.305	106	1050	0.687	ug/l	74
84) 1,2,4-Trimethylbenzene	11.756	105	1821	0.652	ug/l	98
95) Naphthalene	13.780	128	2474	2.653	ug/l	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
Data File : VX023012.D
Acq On : 29 Jun 2021 18:40
Operator : JC/MD
Sample : M2889-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-2

Quant Time: Jun 30 01:45:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

