

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023013.D
 Acq On : 29 Jun 2021 19:04
 Operator : JC/MD
 Sample : M2889-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 TW-3

Quant Time: Jun 30 01:45:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54493	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112336	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36214	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53521	54.120	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.240%
35) Dibromofluoromethane	5.391	113	35019	49.090	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	8.653	98	141729	50.694	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.380%
62) 4-Bromofluorobenzene	11.085	95	48076	46.076	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.160%
Target Compounds						
						Qvalue
3) Chloromethane	1.294	50	312	0.370	ug/l	98
16) Acetone	2.392	43	1651	3.766	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	1245	0.492	ug/l	99
95) Naphthalene	13.780	128	1357	2.365	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023013.D
 Acq On : 29 Jun 2021 19:04
 Operator : JC/MD
 Sample : M2889-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 TW-3

Quant Time: Jun 30 01:45:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

