

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023014.D
 Acq On : 29 Jun 2021 19:27
 Operator : JC/MD
 Sample : M2889-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-1

Quant Time: Jun 30 01:46:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56645	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	119062	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	103182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54340	52.861	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.720%
35) Dibromofluoromethane	5.397	113	37226	49.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	8.653	98	147283	49.704	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.400%
62) 4-Bromofluorobenzene	11.085	95	50556	45.715	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.440%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1959	4.299	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	841	0.308	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
Data File : VX023014.D
Acq On : 29 Jun 2021 19:27
Operator : JC/MD
Sample : M2889-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
TW-1

Quant Time: Jun 30 01:46:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

