

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
 Data File : VX023016.D
 Acq On : 29 Jun 2021 20:14
 Operator : JC/MD
 Sample : M2875-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Jun 30 01:46:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57359	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	119789	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53946	51.824	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	=	103.640%	
35) Dibromofluoromethane	5.397	113	36940	48.561	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	97.120%	
50) Toluene-d8	8.653	98	149170	50.036	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	=	100.080%	
62) 4-Bromofluorobenzene	11.085	95	50219	45.135	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	=	90.260%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3280	7.108	ug/l	100
20) Methylene Chloride	2.794	84	1883	2.414	ug/l	# 79
52) Toluene	8.720	92	2361	1.065	ug/l	88
68) m/p-Xylenes	10.305	106	1128	0.731	ug/l	91
84) 1,2,4-Trimethylbenzene	11.750	105	1221	0.456	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062921\
Data File : VX023016.D
Acq On : 29 Jun 2021 20:14
Operator : JC/MD
Sample : M2875-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Jun 30 01:46:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

