

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029850.D
 Acq On : 29 Jun 2022 10:11
 Operator : JC/MD
 Sample : VX0629WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0629WBL01

Quant Time: Jun 29 14:58:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	254871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	434488	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136457	42.296	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.600%
35) Dibromofluoromethane	5.391	113	124704	45.926	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.860%
50) Toluene-d8	8.653	98	494542	48.549	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.100%
62) 4-Bromofluorobenzene	11.079	95	166137	46.139	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029850.D
Acq On : 29 Jun 2022 10:11
Operator : JC/MD
Sample : VX0629WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0629WBL01

Quant Time: Jun 29 14:58:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

