

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029853.D
 Acq On : 29 Jun 2022 11:52
 Operator : JC/MD
 Sample : N3494-14DL 10X
 Misc : 5.81g/5.00mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-EXC-1-5-GRDL

Quant Time: Jun 29 15:01:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	247637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	419535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362309	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133772	42.675	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.360%
35) Dibromofluoromethane	5.385	113	118159	45.067	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.140%
50) Toluene-d8	8.653	98	479638	48.764	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	161183	46.359	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.720%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.465	56	4571	1.046	ug/l	89
40) Benzene	6.038	78	4049	0.359	ug/l #	86
65) Chlorobenzene	10.080	112	35783	4.920	ug/l	99
70) Styrene	10.659	104	17130	2.109	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	122839	23.440	ug/l	99
95) Naphthalene	13.780	128	8368	0.649	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029853.D
Acq On : 29 Jun 2022 11:52
Operator : JC/MD
Sample : N3494-14DL 10X
Misc : 5.81g/5.00mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-EXC-1-5-GRDL

Quant Time: Jun 29 15:01:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

