

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029855.D
 Acq On : 29 Jun 2022 12:39
 Operator : JC/MD
 Sample : N3494-29DL 10X
 Misc : 5.60g/5.00mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-EXC-1-2-GRDL

Quant Time: Jun 29 15:02:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	248603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	427003	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133917	42.555	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.120%
35) Dibromofluoromethane	5.391	113	118553	44.426	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.860%
50) Toluene-d8	8.653	98	485338	48.481	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.960%
62) 4-Bromofluorobenzene	11.085	95	159839	45.168	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.340%
Target Compounds						
65) Chlorobenzene	10.079	112	344612	46.734	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029855.D
Acq On : 29 Jun 2022 12:39
Operator : JC/MD
Sample : N3494-29DL 10X
Misc : 5.60g/5.00mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-EXC-1-2-GRDL

Quant Time: Jun 29 15:02:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

