

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029861.D
 Acq On : 29 Jun 2022 14:56
 Operator : JC/MD
 Sample : N3505-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1-20220624

Quant Time: Jun 29 22:59:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	241665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	421794	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133774	43.730	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.460%
35) Dibromofluoromethane	5.391	113	116853	44.330	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.660%
50) Toluene-d8	8.653	98	471887	47.719	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.440%
62) 4-Bromofluorobenzene	11.085	95	158945	45.470	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.940%
Target Compounds						
					Qvalue	
30) Chloroform	5.099	83	5198	1.083	ug/l	93
44) Trichloroethene	7.135	130	1035	0.339	ug/l	89
64) Tetrachloroethene	9.275	164	927	0.429	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029861.D
Acq On : 29 Jun 2022 14:56
Operator : JC/MD
Sample : N3505-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1-20220624

Quant Time: Jun 29 22:59:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

