

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029862.D
 Acq On : 29 Jun 2022 15:20
 Operator : JC/MD
 Sample : N3505-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20220624

Quant Time: Jun 29 22:59:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	248916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	432175	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136272	43.249	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.500%
35) Dibromofluoromethane	5.391	113	122851	45.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.980%
50) Toluene-d8	8.653	98	488928	48.255	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.500%
62) 4-Bromofluorobenzene	11.085	95	167041	46.639	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.280%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3241	2.103	ug/l #	85
20) Methylene Chloride	2.782	84	755	0.232	ug/l #	73
44) Trichloroethene	7.129	130	1015	0.324	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029862.D
Acq On : 29 Jun 2022 15:20
Operator : JC/MD
Sample : N3505-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20220624

Quant Time: Jun 29 22:59:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

