

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029869.D
 Acq On : 29 Jun 2022 18:04
 Operator : JC/MD
 Sample : N3485-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GM38-RW3-MW1-20220623

Quant Time: Jun 29 23:01:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	225891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	389490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	330460	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	125298	43.820	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.640%
35) Dibromofluoromethane	5.391	113	111639	45.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.720%
50) Toluene-d8	8.653	98	440047	48.190	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.380%
62) 4-Bromofluorobenzene	11.085	95	143128	44.342	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.680%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2905	2.077	ug/l	91
44) Trichloroethene	7.135	130	62111	22.031	ug/l	93
64) Tetrachloroethene	9.281	164	3837	1.925	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029869.D
Acq On : 29 Jun 2022 18:04
Operator : JC/MD
Sample : N3485-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GM38-RW3-MW1-20220623

Quant Time: Jun 29 23:01:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

