

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029870.D
 Acq On : 29 Jun 2022 18:27
 Operator : JC/MD
 Sample : N3485-02 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20220622

Quant Time: Jun 29 23:01:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	402430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	342511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146521	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	128963	44.673	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.340%		
35) Dibromofluoromethane	5.391	113	114864	45.672	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.340%		
50) Toluene-d8	8.653	98	455083	48.234	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.460%		
62) 4-Bromofluorobenzene	11.085	95	150890	45.243	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.480%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	3937	1.721	ug/l	97
12) 1,1-Dichloroethene	2.319	96	1413	0.602	ug/l #	77
20) Methylene Chloride	2.794	84	793	0.266	ug/l	86
44) Trichloroethene	7.129	130	235861	80.972	ug/l	99
64) Tetrachloroethene	9.281	164	644	0.312	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029870.D
Acq On : 29 Jun 2022 18:27
Operator : JC/MD
Sample : N3485-02 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20220622

Quant Time: Jun 29 23:01:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

