

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029871.D
 Acq On : 29 Jun 2022 18:50
 Operator : JC/MD
 Sample : N3485-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20220623DL

Quant Time: Jun 29 23:01:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	230474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	396389	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	338578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	125821	43.128	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.260%		
35) Dibromofluoromethane	5.391	113	113936	45.993	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.980%		
50) Toluene-d8	8.653	98	449265	48.343	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.680%		
62) 4-Bromofluorobenzene	11.085	95	150419	45.789	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.580%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	941	0.407	ug/l #	87
16) Acetone	2.386	43	3983	2.791	ug/l #	84
44) Trichloroethene	7.129	130	38281	13.342	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029871.D
Acq On : 29 Jun 2022 18:50
Operator : JC/MD
Sample : N3485-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20220623DL

Quant Time: Jun 29 23:01:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

