

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029872.D
 Acq On : 29 Jun 2022 19:13
 Operator : JC/MD
 Sample : N3485-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20220623

Quant Time: Jun 29 23:01:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	228880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	393457	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	339141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127088	43.865	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.740%
35) Dibromofluoromethane	5.391	113	111240	45.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.480%
50) Toluene-d8	8.653	98	449611	48.741	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.480%
62) 4-Bromofluorobenzene	11.085	95	149186	45.752	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.500%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	5808	2.530	ug/l	96
12) 1,1-Dichloroethene	2.319	96	695	0.295	ug/l #	76
16) Acetone	2.386	43	8154	5.753	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	1120	0.367	ug/l	86
44) Trichloroethene	7.129	130	192710	67.667	ug/l	99
64) Tetrachloroethene	9.275	164	1107	0.541	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029872.D
Acq On : 29 Jun 2022 19:13
Operator : JC/MD
Sample : N3485-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20220623

Quant Time: Jun 29 23:01:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

