

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029877.D
 Acq On : 29 Jun 2022 22:14
 Operator : JC/MD
 Sample : VX0629WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0629WBL02

Quant Time: Jun 29 23:03:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	235157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	412003	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	355782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153635	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	130571	43.865	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.720%
35) Dibromofluoromethane	5.391	113	116017	45.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.120%
50) Toluene-d8	8.653	98	465197	48.160	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	11.085	95	157500	46.128	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029877.D
Acq On : 29 Jun 2022 22:14
Operator : JC/MD
Sample : VX0629WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0629WBL02

Quant Time: Jun 29 23:03:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

