

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029880.D
 Acq On : 29 Jun 2022 23:47
 Operator : JC/MD
 Sample : N3481-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Jun 30 06:25:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	225156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	380649	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	322988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	121051	42.473	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.940%
35) Dibromofluoromethane	5.391	113	110235	46.339	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.680%
50) Toluene-d8	8.653	98	431738	48.378	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.760%
62) 4-Bromofluorobenzene	11.085	95	139370	44.180	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.360%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3204	2.298	ug/l	91
17) Carbon Disulfide	2.514	76	2305	0.400	ug/l #	93
29) Tetrahydrofuran	5.032	42	5457	3.426	ug/l	96
44) Trichloroethene	7.129	130	1115	0.405	ug/l	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029880.D
Acq On : 29 Jun 2022 23:47
Operator : JC/MD
Sample : N3481-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Jun 30 06:25:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

