

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029885.D
 Acq On : 30 Jun 2022 01:44
 Operator : JC/MD
 Sample : N3481-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18

Quant Time: Jun 30 06:26:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	210973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	372503	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	318905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	121834	45.621	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.240%
35) Dibromofluoromethane	5.391	113	106710	45.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.680%
50) Toluene-d8	8.653	98	420937	48.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.400%
62) 4-Bromofluorobenzene	11.085	95	140553	45.530	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.060%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8708	6.665	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	8478	1.156	ug/l	96
31) Cyclohexane	5.464	56	4974	1.336	ug/l #	85
39) Methylcyclohexane	7.385	83	3195	0.853	ug/l #	82
44) Trichloroethene	7.129	130	911	0.338	ug/l	57
78) n-propylbenzene	11.311	91	2903	0.245	ug/l	93
95) Naphthalene	13.780	128	3447	0.308	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029885.D
Acq On : 30 Jun 2022 01:44
Operator : JC/MD
Sample : N3481-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18

Quant Time: Jun 30 06:26:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

