

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029888.D
 Acq On : 30 Jun 2022 02:54
 Operator : JC/MD
 Sample : N3481-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-21R

Quant Time: Jun 30 06:27:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	403314	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	347750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	129270	43.985	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.980%
35) Dibromofluoromethane	5.391	113	115657	45.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.780%
50) Toluene-d8	8.653	98	457733	48.409	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.820%
62) 4-Bromofluorobenzene	11.085	95	152592	45.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.300%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6199	4.311	ug/l	96
44) Trichloroethene	7.141	130	795	0.272	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029888.D
Acq On : 30 Jun 2022 02:54
Operator : JC/MD
Sample : N3481-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-21R

Quant Time: Jun 30 06:27:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

