

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029889.D
 Acq On : 30 Jun 2022 03:17
 Operator : JC/MD
 Sample : N3481-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-33

Quant Time: Jun 30 06:27:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	233382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	409910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	350049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	130505	44.176	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.360%
35) Dibromofluoromethane	5.391	113	115799	45.204	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.400%
50) Toluene-d8	8.653	98	460210	47.887	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.780%
62) 4-Bromofluorobenzene	11.085	95	152463	44.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.760%
Target Compounds						
44) Trichloroethene	7.129	130	857	0.289	ug/l	Qvalue 66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029889.D
Acq On : 30 Jun 2022 03:17
Operator : JC/MD
Sample : N3481-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-33

Quant Time: Jun 30 06:27:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

