

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029892.D
 Acq On : 30 Jun 2022 04:27
 Operator : JC/MD
 Sample : N3481-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-30D

Quant Time: Jun 30 06:28:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	223160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	385353	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	336594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	125101	44.286	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.580%		
35) Dibromofluoromethane	5.391	113	111402	46.258	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.520%		
50) Toluene-d8	8.653	98	447693	49.554	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.100%		
62) 4-Bromofluorobenzene	11.085	95	147212	46.096	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.200%		
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	6934	9.386	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	34810	4.486	ug/l	94
40) Benzene	6.043	78	3624	0.350	ug/l	94
44) Trichloroethene	7.129	130	806	0.289	ug/l	54

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029892.D
Acq On : 30 Jun 2022 04:27
Operator : JC/MD
Sample : N3481-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-30D

Quant Time: Jun 30 06:28:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

