

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029896.D
 Acq On : 30 Jun 2022 06:00
 Operator : JC/MD
 Sample : N3481-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-26

Quant Time: Jun 30 06:29:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	231388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	400966	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	348638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149606	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	128730	43.950	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.900%
35) Dibromofluoromethane	5.391	113	116264	46.397	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.800%
50) Toluene-d8	8.652	98	455923	48.500	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.000%
62) 4-Bromofluorobenzene	11.085	95	155274	46.728	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.460%
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	145202	18.048	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029896.D
Acq On : 30 Jun 2022 06:00
Operator : JC/MD
Sample : N3481-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-26

Quant Time: Jun 30 06:29:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

