

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029897.D
 Acq On : 30 Jun 2022 06:23
 Operator : JC/MD
 Sample : N3481-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-24

Quant Time: Jun 30 08:53:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	227716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	396225	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	338995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	126653	43.939	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.880%
35) Dibromofluoromethane	5.391	113	113947	46.017	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.040%
50) Toluene-d8	8.653	98	449258	48.362	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.720%
62) 4-Bromofluorobenzene	11.085	95	149720	45.595	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.200%
Target Compounds						
30) Chloroform	5.105	83	2689	0.595	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029897.D
Acq On : 30 Jun 2022 06:23
Operator : JC/MD
Sample : N3481-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-24

Quant Time: Jun 30 08:53:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

