

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX062922\
 Data File : VX029848.D
 Acq On : 29 Jun 2022 08:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 29 14:57:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	122	0.00
2 T	Dichlorodifluoromethane	50.000	53.859	-7.7	120	0.00
3 P	Chloromethane	50.000	50.059	-0.1	113	0.00
4 C	Vinyl Chloride	50.000	52.806	-5.6#	118	0.00
5 T	Bromomethane	50.000	46.419	7.2	113	0.00
6 T	Chloroethane	50.000	43.928	12.1	107	0.00
7 T	Trichlorofluoromethane	50.000	50.756	-1.5	117	0.00
8 T	Diethyl Ether	50.000	49.794	0.4	117	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.158	-0.3	120	0.00
10 T	Methyl Iodide	50.000	46.043	7.9	108	0.00
11 T	Tert butyl alcohol	250.000	219.952	12.0	103	0.01
12 CM	1,1-Dichloroethene	50.000	48.596	2.8#	115	0.00
13 T	Acrolein	250.000	243.056	2.8	112	0.00
14 T	Allyl chloride	50.000	48.490	3.0	113	0.00
15 T	Acrylonitrile	250.000	229.949	8.0	106	0.00
16 T	Acetone	250.000	218.899	12.4	103	0.00
17 T	Carbon Disulfide	50.000	48.314	3.4	111	0.00
18 T	Methyl Acetate	50.000	44.595	10.8	105	0.00
19 T	Methyl tert-butyl Ether	50.000	48.968	2.1	112	0.00
20 T	Methylene Chloride	50.000	45.398	9.2	113	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.851	4.3	114	0.00
22 T	Diisopropyl ether	50.000	48.148	3.7	110	0.00
23 T	Vinyl Acetate	250.000	241.971	3.2	108	0.00
24 P	1,1-Dichloroethane	50.000	48.608	2.8	112	0.00
25 T	2-Butanone	250.000	223.687	10.5	103	0.00
26 T	2,2-Dichloropropane	50.000	53.260	-6.5	122	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.541	0.9	115	0.00
28 T	Bromochloromethane	50.000	49.692	0.6	116	0.00
29 T	Tetrahydrofuran	250.000	228.256	8.7	104	0.00
30 C	Chloroform	50.000	48.962	2.1#	113	-0.01
31 T	Cyclohexane	50.000	51.401	-2.8	120	0.00
32 T	1,1,1-Trichloroethane	50.000	49.168	1.7	115	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.914	6.2	107	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	120	0.00
35 S	Dibromofluoromethane	50.000	47.910	4.2	111	0.00
36 T	1,1-Dichloropropene	50.000	49.334	1.3	115	0.00
37 T	Ethyl Acetate	50.000	45.351	9.3	103	0.00
38 T	Carbon Tetrachloride	50.000	50.473	-0.9	116	0.00
39 T	Methylcyclohexane	50.000	52.469	-4.9	119	0.00
40 TM	Benzene	50.000	49.134	1.7	114	0.00
41 T	Methacrylonitrile	50.000	46.971	6.1	105	0.00
42 TM	1,2-Dichloroethane	50.000	49.134	1.7	110	0.00
43 T	Isopropyl Acetate	50.000	47.808	4.4	108	0.00
44 TM	Trichloroethene	50.000	46.696	6.6	116	0.00
45 C	1,2-Dichloropropane	50.000	48.801	2.4#	112	0.00
46 T	Dibromomethane	50.000	48.917	2.2	111	0.00
47 T	Bromodichloromethane	50.000	49.577	0.8	111	0.00
48 T	Methyl methacrylate	50.000	48.208	3.6	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
 Data File : VX029848.D
 Acq On : 29 Jun 2022 08:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 29 14:57:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	905.873	9.4	102	0.00
50 S	Toluene-d8	50.000	51.024	-2.0	116	0.00
51 T	4-Methyl-2-Pentanone	250.000	234.646	6.1	105	0.00
52 CM	Toluene	50.000	50.449	-0.9#	114	0.00
53 T	t-1,3-Dichloropropene	50.000	52.235	-4.5	113	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.300	-2.6	113	0.00
55 T	1,1,2-Trichloroethane	50.000	48.016	4.0	110	0.00
56 T	Ethyl methacrylate	50.000	49.465	1.1	109	0.00
57 T	1,3-Dichloropropane	50.000	48.567	2.9	111	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	229.877	8.0	104	0.00
59 T	2-Hexanone	250.000	233.155	6.7	104	0.00
60 T	Dibromochloromethane	50.000	49.803	0.4	111	0.00
61 T	1,2-Dibromoethane	50.000	48.512	3.0	111	0.00
62 S	4-Bromofluorobenzene	50.000	51.716	-3.4	115	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	119	0.00
64 T	Tetrachloroethene	50.000	51.676	-3.4	125	0.00
65 PM	Chlorobenzene	50.000	49.868	0.3	117	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.446	1.1	109	0.00
67 C	Ethyl Benzene	50.000	50.586	-1.2#	117	0.00
68 T	m/p-Xylenes	100.000	101.389	-1.4	117	0.00
69 T	o-Xylene	50.000	49.670	0.7	114	0.00
70 T	Styrene	50.000	50.875	-1.8	116	0.00
71 P	Bromoform	50.000	51.723	-3.4	109	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	124	0.00
73 T	Isopropylbenzene	50.000	47.958	4.1	117	0.00
74 T	N-amyl acetate	50.000	47.482	5.0	107	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	44.225	11.5	107	0.00
76 T	1,2,3-Trichloropropane	50.000	46.560	6.9	109	0.00
77 T	Bromobenzene	50.000	46.842	6.3	113	0.00
78 T	n-propylbenzene	50.000	48.988	2.0	117	0.00
79 T	2-Chlorotoluene	50.000	47.824	4.4	115	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.597	4.8	115	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.563	-3.1	111	0.00
82 T	4-Chlorotoluene	50.000	47.722	4.6	117	0.00
83 T	tert-Butylbenzene	50.000	46.963	6.1	113	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.999	4.0	115	0.00
85 T	sec-Butylbenzene	50.000	48.662	2.7	117	0.00
86 T	p-Isopropyltoluene	50.000	50.775	-1.5	120	0.00
87 T	1,3-Dichlorobenzene	50.000	49.259	1.5	122	0.00
88 T	1,4-Dichlorobenzene	50.000	47.881	4.2	120	0.00
89 T	n-Butylbenzene	50.000	51.773	-3.5	124	0.00
90 T	Hexachloroethane	50.000	51.109	-2.2	121	0.00
91 T	1,2-Dichlorobenzene	50.000	49.045	1.9	119	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.184	1.6	117	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.210	3.6	117	0.00
94 T	Hexachlorobutadiene	50.000	46.789	6.4	118	0.00
95 T	Naphthalene	50.000	48.684	2.6	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062922\
Data File : VX029848.D
Acq On : 29 Jun 2022 08:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 29 14:57:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.299	3.4	116	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6