

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062923\
 Data File : VX036368.D
 Acq On : 29 Jun 2023 11:00
 Operator : JC/MD
 Sample : VX0629WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0629WBL01

Quant Time: Jun 30 01:27:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	303408	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	496866	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	442509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	205697	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	185862	42.735	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.480%
35) Dibromofluoromethane	5.385	113	156103	47.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.300%
50) Toluene-d8	8.647	98	583999	47.726	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.460%
62) 4-Bromofluorobenzene	11.079	95	213382	49.507	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062923\
Data File : VX036368.D
Acq On : 29 Jun 2023 11:00
Operator : JC/MD
Sample : VX0629WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0629WBL01

Quant Time: Jun 30 01:27:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

