

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062923\
 Data File : VX036371.D
 Acq On : 29 Jun 2023 12:26
 Operator : JC/MD
 Sample : 03451-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jun 30 01:29:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	287519	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	475917	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	414438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	188557	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	179037	43.441	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.880%
35) Dibromofluoromethane	5.385	113	148016	47.168	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.340%
50) Toluene-d8	8.647	98	558911	47.686	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.380%
62) 4-Bromofluorobenzene	11.079	95	199513	48.327	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.660%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	17228	3.337	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	118200	30.227	ug/l	83
44) Trichloroethene	7.123	130	18073	4.973	ug/l	92
64) Tetrachloroethene	9.275	164	32192	11.001	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062923\
Data File : VX036371.D
Acq On : 29 Jun 2023 12:26
Operator : JC/MD
Sample : 03451-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jun 30 01:29:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

