

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062923\
 Data File : VX036378.D
 Acq On : 29 Jun 2023 15:11
 Operator : JC/MD
 Sample : 03451-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5RE

Quant Time: Jun 30 01:31:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	205892	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	342399	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	300841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	130009	44.051	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.100%
35) Dibromofluoromethane	5.385	113	107757	47.729	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	8.646	98	404448	47.963	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.920%
62) 4-Bromofluorobenzene	11.079	95	138940	46.778	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%
Target Compounds						
27) cis-1,2-Dichloroethene	4.495	96	5001	1.786	ug/l	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062923\
Data File : VX036378.D
Acq On : 29 Jun 2023 15:11
Operator : JC/MD
Sample : 03451-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5RE

Quant Time: Jun 30 01:31:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

