

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063018\
 Data File : VX003074.D
 Acq On : 30 Jun 2018 13:10
 Operator : JC/MD
 Sample : J3668-03DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021RW245-180620DL

Quant Time: Jul 02 03:19:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	298198	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	413991	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	388919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	222812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	192373	46.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.34%	
35) Dibromofluoromethane	5.51	113	157779	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
50) Toluene-d8	8.72	98	583332	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
62) 4-Bromofluorobenzene	11.14	95	208078	45.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.14%	
Target Compounds						
4) Vinyl Chloride	1.40	62	57396	17.02	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	228221	63.55	ug/l	82
65) Chlorobenzene	10.14	112	15789	1.60	ug/l	100
91) 1,2-Dichlorobenzene	12.40	146	1878	0.24	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063018\
Data File : VX003074.D
Acq On : 30 Jun 2018 13:10
Operator : JC/MD
Sample : J3668-03DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021RW245-180620DL

Quant Time: Jul 02 03:19:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

