

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063018\
 Data File : VX003080.D
 Acq On : 30 Jun 2018 15:50
 Operator : JC/MD
 Sample : J3675-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-0221

Manual Integrations
 APPROVED

apatel
 7/2/2018 4:19:15 PM

Quant Time: Jul 02 03:37:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269756	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	375690	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	349883	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	202991	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	175107	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
35) Dibromofluoromethane	5.51	113	144608	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
50) Toluene-d8	8.72	98	528021	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.14	95	186577	44.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	66811	21.90	ug/l #	76
10) Methyl Iodide	2.52	142	1609	6.69	ug/l	94
16) Acetone	2.45	43	3240	2.10	ug/l	96
21) trans-1,2-Dichloroethene	3.17	96	2282	0.79	ug/l #	71
27) cis-1,2-Dichloroethene	4.60	96	11249	3.46	ug/l	92
31) Cyclohexane	5.57	56	2489	0.53	ug/l #	84
39) Methylcyclohexane	7.46	83	1777	0.37	ug/l	95
52) Toluene	8.79	92	1988	0.26	ug/l	94
67) Ethyl Benzene	10.26	91	3602	0.25	ug/l	92
73) Isopropylbenzene	11.02	105	3812	0.28	ug/l	95
83) tert-Butylbenzene	11.77	119	2989	0.26	ug/l	86
85) sec-Butylbenzene	11.95	105	21757	1.58	ug/l	84
89) n-Butylbenzene	12.39	91	2456m	0.23	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063018\
Data File : VX003080.D
Acq On : 30 Jun 2018 15:50
Operator : JC/MD
Sample : J3675-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S73-0221

Manual Integrations
APPROVED

apatel
7/2/2018 4:19:15 PM

Quant Time: Jul 02 03:37:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

