

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063018\
 Data File : VX003083.D
 Acq On : 30 Jun 2018 17:10
 Operator : JC/MD
 Sample : J3716-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S22-0038

Quant Time: Jul 02 03:46:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	265286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	369621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	344403	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	194738	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	173303	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
35) Dibromofluoromethane	5.51	113	142917	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
50) Toluene-d8	8.72	98	515162	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	11.14	95	181903	44.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.26%	
Target Compounds						
16) Acetone	2.45	43	4842	3.19	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063018\
Data File : VX003083.D
Acq On : 30 Jun 2018 17:10
Operator : JC/MD
Sample : J3716-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S22-0038

Quant Time: Jul 02 03:46:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

