

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063018\
 Data File : VX003103.D
 Acq On : 01 Jul 2018 03:20
 Operator : JC/MD
 Sample : J3800-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ULMER-PARK-FIELD-BLANK-1-6-26

Quant Time: Jul 03 11:21:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	280344	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	390590	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	360245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	207859	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	178050	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
35) Dibromofluoromethane	5.51	113	146618	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	8.72	98	541480	47.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.88%	
62) 4-Bromofluorobenzene	11.14	95	193279	44.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.74%	
Target Compounds						
20) Methylene Chloride	2.86	84	33286	10.66	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063018\
Data File : VX003103.D
Acq On : 01 Jul 2018 03:20
Operator : JC/MD
Sample : J3800-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ULMER-PARK-FIELD-BLANK-1-6-26

Quant Time: Jul 03 11:21:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

