

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063018\
 Data File : VX003107.D
 Acq On : 01 Jul 2018 05:07
 Operator : JC/MD
 Sample : J3800-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: Jul 03 11:27:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	299314	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	406065	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	376762	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	218073	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	188781	45.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.26%	
35) Dibromofluoromethane	5.51	113	161348	50.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.36%	
50) Toluene-d8	8.72	98	565143	48.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.26%	
62) 4-Bromofluorobenzene	11.14	95	203608	44.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.92%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063018\
Data File : VX003107.D
Acq On : 01 Jul 2018 05:07
Operator : JC/MD
Sample : J3800-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Jul 03 11:27:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

