

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063018\
 Data File : VX003108.D
 Acq On : 01 Jul 2018 05:33
 Operator : JC/MD
 Sample : J3800-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15

Quant Time: Jul 02 05:38:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	273922	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	385041	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	358606	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204352	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	180561	47.69	ug/l	0.00
Spiked Amount			Recovery	=	95.38%	
35) Dibromofluoromethane	5.51	113	147560	48.39	ug/l	0.00
Spiked Amount			Recovery	=	96.78%	
50) Toluene-d8	8.72	98	534737	48.02	ug/l	0.00
Spiked Amount			Recovery	=	96.04%	
62) 4-Bromofluorobenzene	11.14	95	190338	44.32	ug/l	0.00
Spiked Amount			Recovery	=	88.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	8719	5.56	ug/l	94
19) Methyl tert-butyl Ether	3.21	73	10979	1.13	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063018\
Data File : VX003108.D
Acq On : 01 Jul 2018 05:33
Operator : JC/MD
Sample : J3800-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15

Quant Time: Jul 02 05:38:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

