

Data File : VX003109.D
Acq On : 01 Jul 2018 06:00
Operator : JC/MD
Sample : J3800-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19

Quant Time: Jul 09 04:00:52 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	287642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	398668	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	377780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	220816	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	187528	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
35) Dibromofluoromethane	5.51	113	155829	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	8.72	98	562645	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.14	95	203436	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
Target Compounds						
16) Acetone	2.45	43	19734	11.985	ug/l	97
25) 2-Butanone	4.71	43	21503	8.800	ug/l	98
30) Chloroform	5.21	83	12913	2.166	ug/l	99
51) 4-Methyl-2-Pentanone	8.65	43	31916	6.462	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX003109.D

Acq On : 01 Jul 2018 06:00

Operator : JC/MD

Sample : J3800-10

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 43 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW-19

Quant Time: Jul 09 04:00:52 2018

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M

Quant Title : SW846 8260

QLast Update : Mon Jun 25 14:38:05 2018

Response via : Initial Calibration

