

Data File : VX003116.D
Acq On : 01 Jul 2018 09:05
Operator : JC/MD
Sample : J3799-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Jul 07 05:17:11 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	283653	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	394411	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	370934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	218128	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	183695	46.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.70%	
35) Dibromofluoromethane	5.51	113	151487	48.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.00%	
50) Toluene-d8	8.72	98	554681	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
62) 4-Bromofluorobenzene	11.14	95	203354	46.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.46%	

Target Compounds				Qvalue		
16) Acetone	2.45	43	10477	6.453	ug/l	100
19) Methyl tert-butyl Ether	3.21	73	3166	0.315	ug/l #	80
39) Methylcyclohexane	7.46	83	2120	0.421	ug/l	98
73) Isopropylbenzene	11.02	105	31453	2.143	ug/l	99
78) n-propylbenzene	11.36	91	36110	2.162	ug/l	97
83) tert-Butylbenzene	11.77	119	9279	0.762	ug/l	87
84) 1,2,4-Trimethylbenzene	11.81	105	8470	0.661	ug/l	96
85) sec-Butylbenzene	11.95	105	33246	2.250	ug/l	93
89) n-Butylbenzene	12.39	91	5995	0.518	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX003116.D
Acq On : 01 Jul 2018 09:05
Operator : JC/MD
Sample : J3799-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Jul 07 05:17:11 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

