

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX063018\
 Data File : VX003094.D
 Acq On : 30 Jun 2018 22:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 02 04:33:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.497	0.519	-4.4	121	0.00
3 P	Chloromethane	0.521	0.494	5.2	109	0.00
4 C	Vinyl Chloride	0.565	0.568	-0.5#	113	0.00
5 T	Bromomethane	0.349	0.209	40.1#	85	0.00
6 T	Chloroethane	0.376	0.352	6.4	107	0.00
7 T	Trichlorofluoromethane	0.954	0.886	7.1	99	0.00
8 T	Diethyl Ether	0.347	0.335	3.5	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.553	0.511	7.6	102	0.00
10 T	Methyl Iodide	0.515	0.516	-0.2	106	0.00
11 T	Tert butyl alcohol	0.138	0.127	8.0	97	0.00
12 CM	1,1-Dichloroethene	0.493	0.466	5.5#	104	0.00
13 T	Acrolein	0.082	0.045	45.1#	63	0.00
14 T	Allyl chloride	0.972	0.906	6.8	100	0.00
15 T	Acrylonitrile	0.294	0.289	1.7	104	0.00
16 T	Acetone	0.286	0.265	7.3	98	0.00
17 T	Carbon Disulfide	1.314	1.210	7.9	103	0.00
18 T	Methyl Acetate	0.800	0.756	5.5	94	0.00
19 T	Methyl tert-butyl Ether	1.773	1.700	4.1	101	0.00
20 T	Methylene Chloride	0.557	0.531	4.7	103	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.503	6.3	103	0.00
22 T	Diisopropyl ether	1.858	1.830	1.5	105	0.00
23 T	Vinyl Acetate	1.573	1.552	1.3	105	0.00
24 P	1,1-Dichloroethane	1.010	0.957	5.2	102	0.00
25 T	2-Butanone	0.425	0.431	-1.4	106	0.00
26 T	2,2-Dichloropropane	0.821	0.566	31.1#	73	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.607	-0.8	109	0.00
28 T	Bromochloromethane	0.483	0.471	2.5	100	0.00
29 T	Tetrahydrofuran	0.272	0.274	-0.7	106	0.00
30 C	Chloroform	1.036	1.015	2.0#	104	0.00
31 T	Cyclohexane	0.867	0.834	3.8	102	0.00
32 T	1,1,1-Trichloroethane	0.912	0.906	0.7	104	0.00
33 S	1,2-Dichloroethane-d4	0.691	0.661	4.3	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.396	0.405	-2.3	106	0.00
36 T	1,1-Dichloropropene	0.548	0.528	3.6	102	0.00
37 T	Ethyl Acetate	0.590	0.611	-3.6	108	0.00
38 T	Carbon Tetrachloride	0.602	0.590	2.0	102	0.00
39 T	Methylcyclohexane	0.639	0.592	7.4	97	0.00
40 TM	Benzene	1.593	1.591	0.1	104	0.00
41 T	Methacrylonitrile	0.340	0.353	-3.8	107	0.00
42 TM	1,2-Dichloroethane	0.644	0.609	5.4	100	0.00
43 T	Isopropyl Acetate	0.990	0.984	0.6	104	0.00
44 TM	Trichloroethene	0.475	0.465	2.1	105	0.00
45 C	1,2-Dichloropropane	0.423	0.419	0.9#	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX063018\
 Data File : VX003094.D
 Acq On : 30 Jun 2018 22:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 02 04:33:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.293	0.288	1.7	102	0.00
47 T	Bromodichloromethane	0.542	0.548	-1.1	103	0.00
48 T	Methyl methacrylate	0.510	0.508	0.4	101	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	108	0.00
50 S	Toluene-d8	1.446	1.431	1.0	101	0.00
51 T	4-Methyl-2-Pentanone	0.619	0.637	-2.9	104	0.00
52 CM	Toluene	1.021	1.029	-0.8#	103	0.00
53 T	t-1,3-Dichloropropene	0.614	0.623	-1.5	100	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.572	3.2	97	0.00
55 T	1,1,2-Trichloroethane	0.419	0.427	-1.9	106	0.00
56 T	Ethyl methacrylate	0.613	0.650	-6.0	107	0.00
57 T	1,3-Dichloropropane	0.689	0.696	-1.0	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.296	0.304	-2.7	102	0.00
59 T	2-Hexanone	0.471	0.485	-3.0	102	0.00
60 T	Dibromochloromethane	0.441	0.464	-5.2	103	0.00
61 T	1,2-Dibromoethane	0.440	0.451	-2.5	105	0.00
62 S	4-Bromofluorobenzene	0.558	0.561	-0.5	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.556	0.575	-3.4	108	0.00
65 PM	Chlorobenzene	1.270	1.252	1.4	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.461	0.469	-1.7	104	0.00
67 C	Ethyl Benzene	2.098	2.109	-0.5#	103	0.00
68 T	m/p-Xylenes	0.809	0.818	-1.1	102	0.00
69 T	o-Xylene	0.787	0.801	-1.8	103	0.00
70 T	Styrene	1.279	1.343	-5.0	104	0.00
71 P	Bromoform	0.363	0.378	-4.1	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.364	3.456	-2.7	102	0.00
74 T	N-amyl acetate	1.646	1.635	0.7	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.036	1.062	-2.5	107	0.00
76 T	1,2,3-Trichloropropane	0.906	1.034	-14.1	116	0.00
77 T	Bromobenzene	0.958	0.963	-0.5	103	0.00
78 T	n-propylbenzene	3.829	3.848	-0.5	99	0.00
79 T	2-Chlorotoluene	2.318	2.318	0.0	101	0.00
80 T	1,3,5-Trimethylbenzene	2.833	2.887	-1.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.267	0.260	2.6	91	0.00
82 T	4-Chlorotoluene	2.744	2.724	0.7	99	0.00
83 T	tert-Butylbenzene	2.792	2.830	-1.4	100	0.00
84 T	1,2,4-Trimethylbenzene	2.935	2.977	-1.4	100	0.00
85 T	sec-Butylbenzene	3.387	3.372	0.4	98	0.00
86 T	p-Isopropyltoluene	3.045	3.034	0.4	96	0.00
87 T	1,3-Dichlorobenzene	1.749	1.760	-0.6	102	0.00
88 T	1,4-Dichlorobenzene	1.800	1.762	2.1	102	0.00
89 T	n-Butylbenzene	2.650	2.521	4.9	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX063018\
Data File : VX003094.D
Acq On : 30 Jun 2018 22:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 02 04:33:02 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.453	0.449	0.9	96	0.00
91 T	1,2-Dichlorobenzene	1.766	1.759	0.4	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.237	-1.7	99	0.00
93 T	1,2,4-Trichlorobenzene	1.285	1.276	0.7	99	0.00
94 T	Hexachlorobutadiene	0.641	0.622	3.0	97	0.00
95 T	Naphthalene	3.423	3.703	-8.2	104	0.00
96 T	1,2,3-Trichlorobenzene	1.284	1.311	-2.1	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6