

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063018\
 Data File : VX003102.D
 Acq On : 01 Jul 2018 02:54
 Operator : JC/MD
 Sample : J3800-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17

Quant Time: Jul 03 11:25:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	286007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	398679	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	370691	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212539	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	187427	47.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.82%	
35) Dibromofluoromethane	5.51	113	152726	48.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.74%	
50) Toluene-d8	8.72	98	553566	48.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.02%	
62) 4-Bromofluorobenzene	11.14	95	197235	44.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.72%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063018\
Data File : VX003102.D
Acq On : 01 Jul 2018 02:54
Operator : JC/MD
Sample : J3800-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17

Quant Time: Jul 03 11:25:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.01 min

Lab File: VX003102.D

Acq: 01 Jul 2018 02:54

Instrument :

MSVOA_X

ClientSampleId :


MW-17

Tot Ion:168 Resp: 286007

Ion Ratio Lower Upper

168 100

99 48.4 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 47.41 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.01 min

Lab File: VX003102.D


Acq: 01 Jul 2018 02:54

Tgt Ion: 65 Resp: 187427

Ion Ratio Lower Upper

65 100

67 51.2 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

80000


60000


40000

20000

0

Time--> 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX003102.D

Acq: 01 Jul 2018 02:54

Instrument :

MSVOA_X

ClientSampleId :

MW-17


Tot Ion:114 Resp: 398679

Ion Ratio Lower Upper

114 100

63 19.9 0.0 41.4




88 15.3 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.37 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX003102.D

Acq: 01 Jul 2018 02:54


Tgt Ion:113 Resp: 152726

Ion Ratio Lower Upper

113 100

111 102.1 81.4 122.0



192 24.8 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.01 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. 0.00 min
Lab File: VX003102.D
Acq: 01 Jul 2018 02:54




Instrument :
MSVOA_X
ClientSampleId :
MW-17


Tot Ion: 98 Resp: 553566
Ion Ratio Lower Upper
98 100
100 64.1 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 44.36 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX003102.D
Acq: 01 Jul 2018 02:54

Tgt Ion: 95 Resp: 197235
Ion Ratio Lower Upper
95 100
174 98.2 0.0 187.4
176 95.5 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX003102.D
Acq: 01 Jul 2018 02:54

Instrument :
MSVOA_X
ClientSampled :
MW-17

Tot Ion:117 Resp: 370691
Ion Ratio Lower Upper
117 100
82 53.5 45.0 67.4
119 32.2 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. -0.01 min
Lab File: VX003102.D
Acq: 01 Jul 2018 02:54

Tgt Ion:152 Resp: 212539
Ion Ratio Lower Upper
152 100
115 53.4 40.1 120.2
150 157.0 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

