


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX063018\
Data File : VX003095.D
Acq On : 30 Jun 2018 22:28
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Title : SW846 8260
Last Update : Mon Jun 25 14:38:05 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	4530	PASS
75	95	30	60	49.7	11780	PASS
95	95	100	100	100.0	23696	PASS
96	95	5	9	6.6	1574	PASS
173	174	0.00	2	1.2	271	PASS
174	95	50	100	94.2	22312	PASS
175	174	5	9	7.9	1767	PASS
176	174	95	101	97.7	21789	PASS
177	176	5	9	7.7	1673	PASS