

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017033.D
 Acq On : 30 Jun 2020 10:25
 Operator : JC/SP
 Sample : VX0630WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0630WBL01

Quant Time: Jul 01 08:10:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	251682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	427678	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	429136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208384	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	163161	52.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.18%	
35) Dibromofluoromethane	5.46	113	142323	52.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.22%	
50) Toluene-d8	8.70	98	513288	51.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.62%	
62) 4-Bromofluorobenzene	11.12	95	203462	52.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.06%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017033.D
Acq On : 30 Jun 2020 10:25
Operator : JC/SP
Sample : VX0630WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0630WBL01

Quant Time: Jul 01 08:10:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

