

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017042.D
 Acq On : 30 Jun 2020 14:08
 Operator : JC/SP
 Sample : L3147-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB05-20200626

Quant Time: Jul 01 08:35:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	237823	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	404741	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	416760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	216286	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	154445	52.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.36%	
35) Dibromofluoromethane	5.47	113	132680	51.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.64%	
50) Toluene-d8	8.70	98	498270	52.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.26%	
62) 4-Bromofluorobenzene	11.12	95	201969	54.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017042.D
Acq On : 30 Jun 2020 14:08
Operator : JC/SP
Sample : L3147-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB05-20200626

Quant Time: Jul 01 08:35:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

