

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017043.D
 Acq On : 30 Jun 2020 14:31
 Operator : JC/SP
 Sample : L3147-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB04-20200626

Quant Time: Jul 01 08:37:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	231523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	406616	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	420669	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	221479	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	153817	53.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.78%	
35) Dibromofluoromethane	5.46	113	134271	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
50) Toluene-d8	8.70	98	493454	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	11.12	95	206717	55.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.20%	
Target Compounds						
16) Acetone	2.42	43	12176	11.404	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017043.D
Acq On : 30 Jun 2020 14:31
Operator : JC/SP
Sample : L3147-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB04-20200626

Quant Time: Jul 01 08:37:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

