

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX063020\
 Data File : VX017050.D
 Acq On : 30 Jun 2020 17:10
 Operator : JC/SP
 Sample : L3123-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jul 05 23:39:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	235914	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	405581	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	411692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	207238	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	153241	52.69	ug/l	0.00
Spiked Amount			Recovery	=	105.38%	
35) Dibromofluoromethane	5.46	113	131050	51.08	ug/l	0.00
Spiked Amount			Recovery	=	102.16%	
50) Toluene-d8	8.70	98	492108	51.87	ug/l	0.00
Spiked Amount			Recovery	=	103.74%	
62) 4-Bromofluorobenzene	11.12	95	198162	53.43	ug/l	0.00
Spiked Amount			Recovery	=	106.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.39	62	69352	27.057	ug/l	98
6) Chloroethane	1.72	64	57200	37.785	ug/l	96
11) Tert butyl alcohol	3.01	59	8206	14.790	ug/l #	94
12) 1,1-Dichloroethene	2.36	96	295122	126.443	ug/l	99
16) Acetone	2.42	43	3818	3.509	ug/l	95
24) 1,1-Dichloroethane	3.67	63	106937	23.541	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	49473	16.857	ug/l	88
39) Methylcyclohexane	7.44	83	2278	0.534	ug/l #	73
40) Benzene	6.12	78	9229	0.819	ug/l	100
42) 1,2-Dichloroethane	6.17	62	25238	6.204	ug/l	97
44) Trichloroethene	7.19	130	3569	0.844	ug/l	95
49) 1,4-Dioxane	7.71	88	169574	2437.815	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX063020\
Data File : VX017050.D
Acq On : 30 Jun 2020 17:10
Operator : JC/SP
Sample : L3123-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jul 05 23:39:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

